


Report Date : 2020/10/21

	Report No. : ECR2035798-1	1 of 2
---	---------------------------	--------

Applicant	SENNO TECHNOLOGY INC.				
Equipment	Sodium Hypochlorite(NaOCl) Meter				
Manufacturer	MIZU	Model	CL-1	Serial No.	EC16091307
Procedure used	GENP-EC-C031(V 1.0)	Received Date	2020/10/7	Calibration Date	2020/10/21
Operator	Sally Liu	Temperature(°C)	(23 ± 3) °C	Relative Humidity(%)	(60 ± 10) %
Add.-Customer Add.-On-Site	516, Sec.4, Junghua Rd., Hsinchu, Taiwan 300, R.O.C.			Location	Laboratory

SGS Standards

Equipment	Manufacturer	Model	Calibration Date
Measuring Cylinder	Witeg	500ml	2020/1/8
Pipette	EPPENDORF	100-1000 μl	2019/11/20
Serial Number	Traceability	Report No.	Due Date
EC15012701	SGS(WUKU)	ECR1944532	2021/1/7
3213346	SGS(TAF 0143)	ECR1937124	2020/11/19

- ◆ SGS Taiwan Ltd hereby declare that traceability follows ILAC P10(ILAC policy on the traceability of measurement results).All standards are directly traceable to TAF recognized lab (members of the ILAC MRA) or to National Metrology Institutes (NMI) or to other international standards (members of the CIPM MRA) or certified reference materials(CRMs). The report with "TAF" accredited symbol represents the certificate of calibration items conforming to TAF accreditation scope ; Without "TAF" accredited symbol , the report also complies with the lab's standard calibration operating procedures and ISO/IEC 17025 requirements.
- ◆ This calibration report is valid only to the items been calibrated. According to ISO / IEC 17025 , SGS will not provide the determination for Calibration interval and acceptable level for this instrument .
- ◆ To reproduce or copy calibration report in partial is not allowed.
- ◆ Procedure used: Calibration SOP for sodium hypochlorite (Bleach) concentration meter

Tina Tang

Approval Signatory

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.



Calibration Result :


Measured Value (ppm)	Actual Value (ppm)	Deviation (ppm)
52	50.0	2.0
199	200.2	-1.2
506	500.4	5.6
999	1000.8	-1.8

Calibration Explanation :

- 1.Deviation = Measured – Actual
 - 2.Deviation % = $[(\text{Measured} - \text{Actual}) \div \text{Actual}] * 100$
 - 3.Measured : The value recorded under testing from item- tested.
 - 4.Actual : The value indicated by traceable standard.
 - 5.The expanded measurement uncertainties estimated are at a level of confidence of approximately 95 % with a coverage factor $k = 2$
 - 6.Expanded uncertainty : 2 ppm
 - 7.The expanded uncertainty included measurement system variances and item-tested variance.
- THE END --

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

Report Date : 2020/10/21


	Report No. : ECR2035800-1	1 of 2
---	---------------------------	--------

Applicant	SENNO TECHNOLOGY INC.				
Equipment	Sodium Hypochlorite(NaOCl) Meter				
Manufacturer	MIZU	Model	CL-4	Serial No.	EC18010301
Procedure used	GENP-EC-C031(V 1.0)	Received Date	2020/10/7	Calibration Date	2020/10/21
Operator	Sally Liu	Temperature(°C)	(23 ± 3) °C	Relative Humidity(%)	(60 ± 10) %
Add.-Customer Add.-On-Site	516, Sec.4, Junghua Rd., Hsinchu, Taiwan 300, R.O.C.			Location	Laboratory

SGS Standards

Equipment	Manufacturer	Model	Calibration Date
Measuring Cylinder	Witeg	500ml	2020/1/8
Pipette	EPPENDORF	100-1000 µl	2019/11/20
Serial Number	Traceability	Report No.	Due Date
EC15012701	SGS(WUKU)	ECR1944532	2021/1/7
3213346	SGS(TAF 0143)	ECR1937124	2020/11/19

- ◆ SGS Taiwan Ltd hereby declare that traceability follows ILAC P10(ILAC policy on the traceability of measurement results).All standards are directly traceable to TAF recognized lab (members of the ILAC MRA) or to National Metrology Institutes (NMI) or to other international standards (members of the CIPM MRA) or certified reference materials(CRMs). The report with "TAF" accredited symbol represents the certificate of calibration items conforming to TAF accreditation scope ; Without "TAF" accredited symbol , the report also complies with the lab's standard calibration operating procedures and ISO/IEC 17025 requirements.
- ◆ This calibration report is valid only to the items been calibrated. According to ISO / IEC 17025 , SGS will not provide the determination for Calibration interval and acceptable level for this instrument .
- ◆ To reproduce or copy calibration report in partial is not allowed.
- ◆ Procedure used: Calibration SOP for sodium hypochlorite (Bleach) concentration meter


 Approval Signatory

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or refer only to the sample(s) tested.



Calibration Result :

Measured Value (%)	Actual Value (%)	Deviation (%)
0.11	0.11	0.00
0.21	0.21	0.00
0.31	0.31	0.00
0.40	0.40	0.00
0.50	0.50	0.00

Calibration Explanation :

- 1.Deviation = Measured – Actual
 - 2.Deviation % = $[(\text{Measured} - \text{Actual}) \div \text{Actual}] * 100$
 - 3.Measured : The value recorded under testing from item- tested.
 - 4.Actual : The value indicated by traceable standard.
 - 5.The expanded measurement uncertainties estimated are at a level of confidence of approximately 95 % with a coverage factor $k = 2$
 - 6.Expanded uncertainty : 0.02 %
 - 7.The expanded uncertainty included measurement system variances and item-tested variance.
- THE END --